



PAPER TESTING ASSOCIATION

CORRUGATED BOARD Testing Equipment




ECT NECK-DOWN CUTTER

Code. 100113

To obtain specimens for ECT test

Standards : TAPPI T838



Sold & serviced by:

OpTest Equipment Inc.

www.optest.com - sales@optest.com - +1-613-632-5169



PAPER TESTING ASSOCIATION

CORRUGATED BOARD Testing Equipment



CHARACTERISTICS :

- Clean and accurate cuts. The sample to be cut, of 2x2 inches, is obtained by means of a previous cut with the ECT Billerud guillotine.
- Robust and easy operating apparatus.
- Cutting of two semicircles with a radius of $R = 12,7$ mm.
- Maximum cutting thickness, 15 mm.
- Samples obtaining regarding to TAPPI T838.
- CE mark.



WEIGHT AND DIMENSIONS (LEVER INCLUDED) :

310 x 100 x 255 mm (WxLxH) / 7 Kg

Sold & serviced by:

OpTest Equipment Inc.

www.optest.com - sales@optest.com - +1-613-632-5169