



Agitation Device for Bathroom Tissue

Code N4300S

A sample of toilette paper is agitated in a define water volume. This apparatus measures the time to desintegrate the sample.

Standards: NFQ 34.020

- Rotor speed: 800 r.p.m \pm 20 r.p.m
- Special head for similar agitation inside the WC.
- Sample: 600 ml.
- 6 sheets of different packages.
- Dimensions of the polytetrafluoroethylene head, 60 ± 1 mm \varnothing and 17 ± 1 mm in height with 4 cross reliefs.
- Dimensions container: Internal \varnothing 98 mm \pm 5 mm, height ≥ 150 mm.
- Test time: up to disintegration.
- Result in seconds (max 2 min).
- Visual appreciation



Sold & serviced by:

OpTest Equipment Inc.

www.optest.com - sales@optest.com - +1-613-632-5169